



## The Model T500U CAPS NO<sub>2</sub> Analyzer



The Model T500U NO<sub>2</sub> analyzer is a significant advancement in the measurement of NO<sub>2</sub> using a Cavity Attenuated Phase Shift (CAPS) spectroscopy technique to provide highly accurate, real-time, continuous, and direct readings.

— Available with NumaView™ premium T Series software —

- Large, vivid, and durable color touchscreen display
- All other T Series instrument platform features
- Lifetime technical support by phone and email
- Standard two-year warranty

# T500U Specifications

■ Ranges	0 - 5 ppb to 0 - 1 ppm NO <sub>2</sub>
■ Measurement Units	ppb, ppm, µg/m <sup>3</sup> , mg/m <sup>3</sup> (selectable)
■ Zero Noise	< 20 ppt (RMS)
■ Span Noise	< 0.2% of reading (RMS) + 20 ppt
■ Lower Detectable Limit	< 40 ppt
■ Zero Drift	< 0.1 ppb / 24 hours
■ Span Drift	< 0.5% of reading / 24 hours
■ Rise/Fall Time	< 30 seconds to 95%
■ Linearity	< 1% of full scale
■ Precision	0.5% of reading above 5 ppb
■ Sample Flow Rate	900 cm <sup>3</sup> /min ±10%
■ Power Requirements	80W; 100-250VAC (50-60Hz)
■ Analog Output Ranges	10V, 5V, 1V, 0.1V (selectable)
■ Recorder Offset	±10%
■ Included I/O	1 x Ethernet: 10/100Base-T 2 x RS232 (300-115,200 baud) 2 x USB device ports 8 x opto-isolated digital outputs 6 x opto-isolated digital inputs 4 x analog outputs
■ Optional I/O	1 x USB com port 1 x RS485 8 x analog inputs (0-10V, 12-bit) 4 x digital alarm outputs Multidrop RS232 3 x 4 - 20mA current outputs
■ Operating Temperature Range	5 - 40°C (with US EPA Approval)
■ Dimensions (HxWxD)	7" x 17" x 23.5" (178 x 432 x 597 mm)
■ Weight	33 lbs (15kg)
■ Certifications*	US EPA: Federal Equivalent Method (EQNA-0514-212) EU: EN14211 TÜV Rheinland QAL1 Certified: EN15267 MCerts: Sira MC160304/00

\* All certifications apply for legacy or NumaView™ T Series analyzer software

NumaView™ software is available as a no-charge option that must be specified at the time of purchase.

Specifications subject to change without notice.  
All specifications are based on constant conditions.